

AN-17-001

25th May 2017

CUSTOMER ADVISORY NOTICE


Subject: This Customer Advisory Notice covers a manufacturing and labelling change for the following product part numbers for INSTI HIV-1/HIV-2 Antibody Test Kit :

- 90-1008
- 90-1012
- 90-1015
- 90-1019
- 90-1021
- 90-1022
- 90-1026

Dear Valued Customer,

Please be advised that a manufacturing and labeling change has been made to the Membrane unit component of the INSTI HIV-1/HIV-2 Antibody Test .

Starting June 1st 2017, all membrane units will be manufactured with "C" print on the membrane unit top cover as displayed in the picture below. This change will assist the user to identify the control dot location and thus improve test result interpretation. This change is an improvement and does not impact the test performance.



For further clarifications or questions, please contact bioLytical Laboratories, at: 1-604-204-6784 or customercare@biolytical.com.

Sincerely,


Angela Tapu
Director, Quality & Regulatory

bioLytical Laboratories

Customer Advisory Notices are confidential and propriety information of bioLytical and is intended only for distribution by bioLytical to its customers, for customers' use only. It must not be copied or provided to any third party without bioLytical's prior written consent.